

Manual Supplement

Manual Title: 345 Users
Part Number: 2560401
Print Date: October 2006
Revision/Date:

Supplement Issue: 1
Issue Date: 2/07
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.



99 Washington Street
Melrose, MA 02176
Phone 781-665-1400
Toll Free 1-800-517-8431



Visit us at www.TestEquipmentDepot.com



[Back to the Fluke 345 Product Info Page](#)

Change #1

On page 15, under ***Voltage Measuring Input***, replace the first sentence with the following:

The maximum input voltage for over-voltage category CAT IV may not exceed 600 V to earth ground.