



Agilent
E363xA Series
Programmable DC Power Supplies

Data Sheet



E3631A/32A/33A/34A Programmable DC Power Supply Specifications

Model Number	E3631A			E3632A	E3633A	E3634A	
	1	2	3				
DC Output Rating (0 °C to 40 °C)	0 to +6 V, 0 to 5 A	0 to +25 V, 0 to 1 A	0 to -25 V, 0 to 1 A	0 to 15 V/7A or 0 to 30 V/4 A	0 to 8 V/20A or 0 to 20 V/10 A	0 to 25 V/7 A or 0 to 50 V/ 4A	
Load Regulation ± (% of output + offset)	< 0.01% + 2 mV < 0.01% + 250 µA						
Line Regulation ± (% of output + offset)	< 0.01% + 2 mV < 0.01% + 250 µA						
Ripple and Noise (20 Hz to 20 MHz)							
Normal Mode Voltage	< 350 µVrms/2 mVpp				< 350 µVrms/3 mVpp	< 500 µVrms/ 3 mVpp	
Normal Mode Current	< 2 mArms	< 500 µArms		< 2 mArms			
Common Mode Current	< 1.5 µArms						
Accuracy¹ 12 Months (25 °C + 5 °C), ± (% output + offset)							
Programming							
Voltage	0.1% + 5 mV	0.05% + 20 mV		0.05% + 10 mV			
Current	0.2% + 10 mA	0.15% + 4 mA		0.2% + 10 mA			
Readback							
Voltage	0.1% + 5 mV	0.05% + 10 mV		0.05% + 5 mV			
Current	0.2% + 10 mA	0.15% + 4 mA		0.15% + 5 mA			
Resolution							
Program	0.5 mV/0.5 mA	1.5 mV/0.1 mA		1 mV/0.5 mA	1 mV/1 mA	3 mV/0.5 mA	
Readback	0.5 mV/0.5 mA	1.5 mV/0.1 mA		0.5 mV/0.1 mA	0.5 mV/1 mA	1.5 mV/0.5 mA	
Meter	1 mV/1 mA	10 mV/1 mA		1 mV/1 mA	1 mV/1mA (< 10A), 10 mA (≥ 10 A)		
Transient Response	Less than 50 µsec for output to recover to within 15 mV following a change in output current from full load to half load or vice versa						
Command Processing Time²	< 50 msec			< 100 msec			
OVP/OCP							
Accuracy ± (% output + offset)	N/A			0.5% + 0.5 V/0.5% + 0.5 A			
Activation Time	N/A			1.5 msec, OVP ≥ 3 V/< 10 msec, OVP < 3 V and OCP <10 msec			
Temperature Coefficient per °C (% output + offset)							
Voltage	0.01% + 2 mV	0.01% + 3 mV					
Current	0.02% + 3 mA	0.02% + 0.5 mA		0.02% + 3 mA			
Stability , constant load & temperature ± (% of output + offset), 8 hrs							
Voltage	0.03% + 1 mV	0.02% + 2 mV		0.02% + 1 mV			
Current	0.1% + 3 mA	0.05% + 1 mA		0.1% + 1 mA			
Remote Sense (max. voltage in each load lead)	N/A			1 V	0.7 V		
Voltage Programming Speed , to within 1% of total excursion							
Up	Full Load	11 msec	50 msec		50 msec	95 msec	80 msec
	No Load	10 msec	20 msec		20 msec	45 msec	100 msec
Down	Full Load	13 msec	45 msec		45 msec	30 msec	30 msec
	No Load	200 msec	400 msec		400 msec	450 msec	450 msec

¹ Accuracy specifications are valid after a 1-hour warm-up and calibration at 25 °C.

² Maximum time for output to change after receipt of commands.

Model Number	E3631A			E3632A	E3633A	E3634A
	1	2	3			
AC Input (47 Hz – 63 Hz)	100 Vac ±10% (Opt 0E9)/115 Vac ±10% (Std)/230 Vac ±10% (0E3)					
Dimensions	213 × mm W × 133 mm H × 348 mm D (8.4 × 5.2 × 13.7 in)					
Weight	8.2 kg (18 lbs) net, 11 kg (24 lbs) shipping			9.5 kg (21 lbs) net, 12 kg (26 lbs) shipping		
Warranty	1 year					
Product Regulation	Certified to CSA 22.2 No. 231 (for E3631A), No. 1010.1 (for E3632A/33A/34A); conforms to IEC 1010-1; carries CE marks; complies with CISPR-11, Group 1, Class A					

Ordering Information

E3630 Series Power Supplies
E3631A 80-Watt
Triple Power Supply
E3632A 120-Watt
Single Power Supply
E3633A/34A 200-Watt
Single Power Supply

Power Options

Opt. 0E3 230 Vac ± 10%
Opt. 0EM 115 Vac ± 10%
Opt. 0E9 100 Vac ± 10%

Other Options

Opt. 0L2 Extra manual sets
Opt. 1CM Rackmount kit*
Opt. UK6 Commercial calibration
with test result data
E3600A-100 Test lead kit

Rackmount Kits*

Agilent E3631A/32A/33A/34A
To rackmount two instruments side-by-side
Lock-link Kit (P/N 5061-9694)
Flange Kit (P/N 5063-9214)
To rackmount one or two instruments in a
sliding support shelf
Support Shelf (P/N 5063-9256)
Slide Kit (P/N 1494-0015) required for
support shelf

* Rackmounting with 1CM or lock-link/
flange kit requires
Agilent or customer support rails
Agilent Support Rails-E3663AC



Agilent Email Updates

www.agilent.com/find/emailupdates

Get the latest information on the products and applications you select.



Agilent Direct

www.agilent.com/find/agilentdirect

Quickly choose and use your test equipment solutions with confidence.



www.agilent.com/find/open

Agilent Open simplifies the process of connecting and programming test systems to help engineers design, validate and manufacture electronic products. Agilent offers open connectivity for a broad range of system-ready instruments, open industry software, PC-standard I/O and global support, which are combined to more easily integrate test system development.



LXI is the US registered trademark of the LXI Consortium.

Remove all doubt

Our repair and calibration services will get your equipment back to you, performing like new, when promised. You will get full value out of your Agilent equipment throughout its lifetime. Your equipment will be serviced by Agilent-trained technicians using the latest factory calibration procedures, automated repair diagnostics and genuine parts. You will always have the utmost confidence in your measurements.

Agilent offers a wide range of additional expert test and measurement services for your equipment, including initial start-up assistance onsite education and training, as well as design, system integration, and project management.

For more information on repair and calibration services, go to

www.agilent.com/find/removealldoubt

www.agilent.com

For more information on Agilent Technologies' products, applications or services, please contact your local Agilent office. The complete list is available at:

www.agilent.com/find/contactus

Phone or Fax

United States:

(tel) 800 829 4444
(fax) 800 829 4433

Canada:

(tel) 877 894 4414
(fax) 800 746 4866

China:

(tel) 800 810 0189
(fax) 800 820 2816

Europe:

(tel) 31 20 547 2111

Japan:

(tel) (81) 426 56 7832
(fax) (81) 426 56 7840

Korea:

(tel) (080) 769 0800
(fax) (080) 769 0900

Latin America:

(tel) (305) 269 7500

Taiwan:

(tel) 0800 047 866
(fax) 0800 286 331

Other Asia Pacific Countries:

(tel) (65) 6375 8100
(fax) (65) 6755 0042
Email: tm_ap@agilent.com
Revised: 1%/06

Product specifications and descriptions in this document subject to change without notice.

© Agilent Technologies, Inc. 2002-2007
Printed in USA, October 26, 2007
5968-9726EN



Agilent Technologies